



Department of Civil Engineering National Institute of Technology, Kurukshetra

No. CED/Envi./NITK/2017/ 26

Dated: - 25.01.2018

M/s. Addressed as per


List attached


SUBJECT: INVITATION OF QUOTATIONS FOR PURCHASE OF JAR TEST APPARATUS.

1. You are invited to submit your most competitive quotation for the following good(s):

Sr. No.	Brief Description & Specification of Goods	Quantity	Delivery Period	F. O. R.
1.	Jar Test Apparatus as per Detailed Specifications on attached separate sheet	01	4-6 Weeks	N.I.T. Kurukshetra

2. Necessary Literature of the goods may please be sent to facilitate to take decision.
3. All duties, taxes and other levies payable by the Institute shall be included in the total price. This Institute does not issue Form C or D.
4. Payment will be made Online through RTGS/NEFT within 30 days after receipt of material in good condition and according to specifications. The bank detail for making online payment may be indicated in the quotation.
5. The item must be supplied within delivery period or delivery period extended by the Institute on the request of the supplier on genuine grounds otherwise the penalty for delayed period @ 0.5% of the amount shall be charged for every week or part thereof and the maximum 10%.
6. The goods are not required exclusively for Research Purpose. The Duties are payable by the Institute.
7. The quotation should remain valid for a period not less than 45 days from the date of submission.
8. The right of accepting or rejecting any quotation and to cancel the bidding process and reject all quotations without assigning any reason is reserved with the Institute.
9. The due date of receipt of Quotation is **15.02.2018** and will be opened on next working day. Please quote on the top of the envelope our Reference No. and due date of opening.


Head of Department (Incharge)
Civil Engineering Department
National Institute of Technology
Kurukshetra-136119


25.01.18
Indentor
Civil Engineering Department
National Institute of Technology
Kurukshetra-136119

Specifications of Jar Test Apparatus

Laboratory microprocessor controlled Jar Test Apparatus complete with all accessories and essential features including robust design, multiple stirrers, LED display and pre-settable programs to allow standard conditions to be met during laboratory study to obtain reliable and repeatable results.

Essential Features:

Construction material: Epoxy painted metal structure

No of Positions: 6

Back/Bottom Panel: Illuminated

Stirring Rods: Stainless steel rods adjustable in height by a self blocking chuck

Motor: Electronically controlled DC gear motor

Power: AC 220 - 240 V, 50-60 Hz

Electronic speed control: 10 to 300 rpm

Speed setting interval: 1 rpm

Microprocessor controlled timer: 0 - 99 minutes or continuous

Warranty: Three years from the date of installation

Installation & Servicing: The equipment should be installed by certified engineer of the firm. The service engineer of the firm shall be visiting as frequently as necessary for repair and maintenance of equipment during 3 year warranty period.

Accessories and Spare Consumables: The instrument should include all essential accessories including operating manual, dust cover, power cord, etc. and spare consumables for satisfactory continuous operation during three year of warranty period